

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072035.D
 Acq On : 21 Feb 2022 16:30
 Operator : VA/SY
 Sample : N1626-06
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B3

Quant Time: Feb 22 01:06:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	111366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	194303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	215739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	105571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	87615	75.844	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.680%
35) Dibromofluoromethane	7.914	113	55291	46.967	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.940%
50) Toluene-d8	10.338	98	240493	55.088	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.180%
62) 4-Bromofluorobenzene	12.626	95	95314	59.437	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.880%
Target Compounds						
16) Acetone	4.156	43	24854	99.552	ug/l	99
25) 2-Butanone	7.214	43	5014	14.531	ug/l #	49

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072035.D
Acq On : 21 Feb 2022 16:30
Operator : VA/SY
Sample : N1626-06
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B3

Quant Time: Feb 22 01:06:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

