

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072037.D
 Acq On : 21 Feb 2022 17:26
 Operator : VA/SY
 Sample : N1626-08
 Misc : 8.13G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B4

Quant Time: Feb 22 01:06:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	351418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	638249	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	646553	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	303988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	235997	64.74	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.48%
35) Dibromofluoromethane	7.914	113	123158	31.85	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	63.70%
50) Toluene-d8	10.337	98	794964	55.44	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.88%
62) 4-Bromofluorobenzene	12.620	95	302781	57.48	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.96%
Target Compounds						
16) Acetone	4.161	43	33808	42.91	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072037.D
Acq On : 21 Feb 2022 17:26
Operator : VA/SY
Sample : N1626-08
Misc : 8.13G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B4

Quant Time: Feb 22 01:06:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

