

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072040.D
 Acq On : 21 Feb 2022 18:51
 Operator : VA/SY
 Sample : N1626-11
 Misc : 6.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A6

Quant Time: Feb 22 01:08:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	447942	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	808416	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	717821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	266804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	246603	53.07	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.14%
35) Dibromofluoromethane	7.914	113	152287	31.09	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	62.18%
50) Toluene-d8	10.338	98	949607	52.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.56%
62) 4-Bromofluorobenzene	12.620	95	314815	47.18	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.36%
Target Compounds						
16) Acetone	4.156	43	72394	72.09	ug/l	97
17) Carbon Disulfide	4.403	76	7383	1.30	ug/l	96
25) 2-Butanone	7.220	43	20286	14.62	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072040.D
Acq On : 21 Feb 2022 18:51
Operator : VA/SY
Sample : N1626-11
Misc : 6.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A6

Quant Time: Feb 22 01:08:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

