

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072041.D
 Acq On : 21 Feb 2022 19:19
 Operator : VA/SY
 Sample : N1626-12
 Misc : 7.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B6

Quant Time: Feb 22 01:09:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	474813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	863910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	820326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	363028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	264100	53.622	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.240%	
35) Dibromofluoromethane	7.914	113	156501	29.900	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 59.800%	
50) Toluene-d8	10.338	98	1061078	54.665	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.340%	
62) 4-Bromofluorobenzene	12.626	95	388835	54.535	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.060%	
Target Compounds						
						Qvalue
16) Acetone	4.162	43	60888	57.203	ug/l	97
17) Carbon Disulfide	4.415	76	9707	1.454	ug/l #	94
25) 2-Butanone	7.220	43	11836	8.045	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072041.D
Acq On : 21 Feb 2022 19:19
Operator : VA/SY
Sample : N1626-12
Misc : 7.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B6

Quant Time: Feb 22 01:09:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

