

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072044.D
 Acq On : 21 Feb 2022 20:44
 Operator : VA/SY
 Sample : N1626-15
 Misc : 6.43G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A8

Quant Time: Feb 22 01:09:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	524958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	914222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	873043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	400940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	264394	48.554	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.100%
35) Dibromofluoromethane	7.914	113	150245	27.125	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	54.240%
50) Toluene-d8	10.338	98	1137140	55.360	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.720%
62) 4-Bromofluorobenzene	12.620	95	412799	54.710	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.420%
Target Compounds						
16) Acetone	4.162	43	18954	16.106	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072044.D
Acq On : 21 Feb 2022 20:44
Operator : VA/SY
Sample : N1626-15
Misc : 6.43G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A8

Quant Time: Feb 22 01:09:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

