

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
 Data File : VD072051.D
 Acq On : 22 Feb 2022 12:17
 Operator : VA/SY
 Sample : N1626-01
 Misc : 4.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A1

Quant Time: Feb 23 05:29:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	564647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	983837	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	935098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	434324	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	280927	47.964	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.920%
35) Dibromofluoromethane	7.915	113	302492	50.747	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.500%
50) Toluene-d8	10.338	98	1206890	54.598	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.200%
62) 4-Bromofluorobenzene	12.626	95	430439	53.011	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
Data File : VD072051.D
Acq On : 22 Feb 2022 12:17
Operator : VA/SY
Sample : N1626-01
Misc : 4.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A1

Quant Time: Feb 23 05:29:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

