

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
 Data File : VD072055.D
 Acq On : 22 Feb 2022 14:11
 Operator : VA/SY
 Sample : N1626-03
 Misc : 5.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A2

Quant Time: Feb 23 05:30:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	545514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	942854	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	888289	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	404169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	269629	47.65	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.30%
35) Dibromofluoromethane	7.914	113	216832	37.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.92%
50) Toluene-d8	10.338	98	1152802	54.42	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.84%
62) 4-Bromofluorobenzene	12.626	95	420385	54.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.04%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	68453	55.97	ug/l	95
25) 2-Butanone	7.208	43	24236	14.34	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
Data File : VD072055.D
Acq On : 22 Feb 2022 14:11
Operator : VA/SY
Sample : N1626-03
Misc : 5.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A2

Quant Time: Feb 23 05:30:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

