

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
 Data File : VD072057.D
 Acq On : 22 Feb 2022 15:07
 Operator : VA/SY
 Sample : N1626-05RE
 Misc : 6.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A3RE

Quant Time: Feb 23 05:30:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	546956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	950986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	905937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	405244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	284750	50.189	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 100.380%		
35) Dibromofluoromethane	7.909	113	146728	25.466	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 50.940%#		
50) Toluene-d8	10.332	98	1174509	54.969	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 109.940%		
62) 4-Bromofluorobenzene	12.620	95	425403	54.201	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 108.400%		
Target Compounds						
						Qvalue
16) Acetone	4.156	43	32438	26.455	ug/l	99
17) Carbon Disulfide	4.409	76	7241	1.175	ug/l	96
93) 1,2,4-Trichlorobenzene	15.143	180	8536	1.203	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
Data File : VD072057.D
Acq On : 22 Feb 2022 15:07
Operator : VA/SY
Sample : N1626-05RE
Misc : 6.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A3RE

Quant Time: Feb 23 05:30:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

