

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
 Data File : VD072061.D
 Acq On : 22 Feb 2022 17:00
 Operator : VA/SY
 Sample : N1626-16
 Misc : 6.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B8

Quant Time: Feb 23 05:31:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	545805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	935297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	859892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	390552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	266746	47.115	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.220%
35) Dibromofluoromethane	7.914	113	272007	48.001	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.000%
50) Toluene-d8	10.338	98	1128658	53.709	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.420%
62) 4-Bromofluorobenzene	12.626	95	406252	52.629	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.260%
Target Compounds						
					Qvalue	
16) Acetone	4.185	43	95356	77.932	ug/l	95
17) Carbon Disulfide	4.403	76	4960	1.014	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022222\
Data File : VD072061.D
Acq On : 22 Feb 2022 17:00
Operator : VA/SY
Sample : N1626-16
Misc : 6.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B8

Quant Time: Feb 23 05:31:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

