

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
 Data File : VD072073.D
 Acq On : 23 Feb 2022 14:38
 Operator : VA/SY
 Sample : N1598-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GS-1

Quant Time: Feb 24 06:46:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	551953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	928091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	887721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	413219	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	256059	44.723	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.440%
35) Dibromofluoromethane	7.908	113	288595	51.324	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.640%
50) Toluene-d8	10.332	98	1150616	55.179	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.360%
62) 4-Bromofluorobenzene	12.620	95	406362	53.052	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
Data File : VD072073.D
Acq On : 23 Feb 2022 14:38
Operator : VA/SY
Sample : N1598-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GS-1

Quant Time: Feb 24 06:46:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

