

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
 Data File : VD072075.D
 Acq On : 23 Feb 2022 15:34
 Operator : VA/SY
 Sample : N1598-03
 Misc : 5.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GS-3

Quant Time: Feb 24 06:47:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	530789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	889067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	842059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	392799	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	244980	44.495	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.980%
35) Dibromofluoromethane	7.914	113	270760	50.266	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.540%
50) Toluene-d8	10.338	98	1107348	55.435	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.860%
62) 4-Bromofluorobenzene	12.626	95	381950	52.053	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
Data File : VD072075.D
Acq On : 23 Feb 2022 15:34
Operator : VA/SY
Sample : N1598-03
Misc : 5.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GS-3

Quant Time: Feb 24 06:47:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

