

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
 Data File : VD072076.D
 Acq On : 23 Feb 2022 16:02
 Operator : VA/SY
 Sample : N1598-04
 Misc : 6.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GS-4

Quant Time: Feb 24 06:47:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	256515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	443721	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	461783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	224471	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	175753	66.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.100%
35) Dibromofluoromethane	7.914	113	155522	57.850	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.700%
50) Toluene-d8	10.338	98	548912	55.059	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.120%
62) 4-Bromofluorobenzene	12.626	95	212784	58.104	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.200%
Target Compounds						
16) Acetone	4.150	43	5912	10.281	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
Data File : VD072076.D
Acq On : 23 Feb 2022 16:02
Operator : VA/SY
Sample : N1598-04
Misc : 6.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GS-4

Quant Time: Feb 24 06:47:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

