

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
 Data File : VD072080.D
 Acq On : 23 Feb 2022 17:55
 Operator : VA/SY
 Sample : N1598-08
 Misc : 6.37G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GS-8

Quant Time: Feb 24 06:47:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	528137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	898402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	846919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	399806	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	251851	45.972	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.940%
35) Dibromofluoromethane	7.914	113	272474	50.058	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.120%
50) Toluene-d8	10.332	98	1094449	54.220	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.440%
62) 4-Bromofluorobenzene	12.626	95	389492	52.530	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022322\
Data File : VD072080.D
Acq On : 23 Feb 2022 17:55
Operator : VA/SY
Sample : N1598-08
Misc : 6.37G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GS-8

Quant Time: Feb 24 06:47:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

