

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030823\
 Data File : VD075471.D
 Acq On : 08 Mar 2023 16:03
 Operator : KP/SY
 Sample : 01833-04
 Misc : 5.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 42

Quant Time: Mar 09 02:16:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 08 12:19:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	204840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	415541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	432985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	203665	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	99029	78.343	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	156.680%
35) Dibromofluoromethane	7.804	113	138376	57.469	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.940%
50) Toluene-d8	10.269	98	530780	60.878	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	121.760%
62) 4-Bromofluorobenzene	12.575	95	168367	63.859	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	127.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030823\
Data File : VD075471.D
Acq On : 08 Mar 2023 16:03
Operator : KP/SY
Sample : 01833-04
Misc : 5.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
42

Quant Time: Mar 09 02:16:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030723S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 08 12:19:57 2023
Response via : Initial Calibration

