

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030823\
 Data File : VD075473.D
 Acq On : 08 Mar 2023 17:01
 Operator : KP/SY
 Sample : 01837-01
 Misc : 13.77G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CB-22-3.5-030723

Quant Time: Mar 09 02:17:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 08 12:19:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	423092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	834623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	761855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	308922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	75998	24.717	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	49.440%#
35) Dibromofluoromethane	7.810	113	124345	25.711	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	51.420%#
50) Toluene-d8	10.275	98	509160	29.075	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	58.160%
62) 4-Bromofluorobenzene	12.575	95	148399	28.023	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	56.040%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.087	96	13637	2.251	ug/l	73
44) Trichloroethene	9.028	130	98697	14.091	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030823\
Data File : VD075473.D
Acq On : 08 Mar 2023 17:01
Operator : KP/SY
Sample : 01837-01
Misc : 13.77G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CB-22-3.5-030723

Quant Time: Mar 09 02:17:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030723S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 08 12:19:57 2023
Response via : Initial Calibration

