

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031424\
 Data File : VD078663.D
 Acq On : 14 Mar 2024 12:51
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 15 05:46:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	99920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	228368	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	307481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	165401	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	45290	48.365	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.740%
35) Dibromofluoromethane	7.804	113	73152	49.752	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.500%
50) Toluene-d8	10.269	98	312148	57.725	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.460%
62) 4-Bromofluorobenzene	12.575	95	128028	80.929	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	161.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031424\
Data File : VD078663.D
Acq On : 14 Mar 2024 12:51
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 15 05:46:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

