

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061201.D
 Acq On : 20 Mar 2019 11:53
 Operator : VA/AP
 Sample : VD0320SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0320SBL01

Quant Time: Mar 22 02:17:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	704623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1174172	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	930133	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	432892	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.45	65	287947	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
35) Dibromofluoromethane	6.91	113	451164	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
50) Toluene-d8	10.40	98	1040865	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
62) 4-Bromofluorobenzene	13.55	95	415058	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061201.D
Acq On : 20 Mar 2019 11:53
Operator : VA/AP
Sample : VD0320SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0320SBL01

Quant Time: Mar 22 02:17:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

