

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061217.D
 Acq On : 20 Mar 2019 19:26
 Operator : VA/AP
 Sample : K2004-03
 Misc : 5.02g/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-19(7.5-8.5)

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 8:56:44 AM

Quant Time: Mar 22 05:08:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	100808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	194096	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	98750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	24819	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	70149	83.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	167.20%	
35) Dibromofluoromethane	6.92	113	85194	59.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.28%	
50) Toluene-d8	10.40	98	171041	47.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.56%	
62) 4-Bromofluorobenzene	13.54	95	301746	214.99	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	429.98%	
Target Compounds						
16) Acetone	3.22	43	50072	208.676	ug/l	98
20) Methylene Chloride	3.82	84	14265	9.236	ug/l	88
25) 2-Butanone	6.02	43	53095	195.003	ug/l #	55
31) Cyclohexane	6.96	56	14588m	9.825	ug/l	
39) Methylcyclohexane	8.83	83	786384	453.286	ug/l #	38
83) tert-Butylbenzene	14.19	119	43394	29.201	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061217.D
Acq On : 20 Mar 2019 19:26
Operator : VA/AP
Sample : K2004-03
Misc : 5.02g/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-19(7.5-8.5)

Manual Integrations
APPROVED

MMDadoda
3/27/2019 8:56:44 AM

Quant Time: Mar 22 05:08:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

