

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061220.D
 Acq On : 20 Mar 2019 20:48
 Operator : VA/AP
 Sample : K2004-07
 Misc : 4.39g/5ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-19B(7-8)

Quant Time: Mar 22 03:29:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	460637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	869036	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	373564	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	52130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	213892	55.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.56%	
35) Dibromofluoromethane	6.92	113	335982	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	10.41	98	577266	35.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.28%	
62) 4-Bromofluorobenzene	13.54	95	455268	72.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	144.90%	
Target Compounds						
					Qvalue	
16) Acetone	3.21	43	73217	66.777	ug/l	98
17) Carbon Disulfide	3.39	76	22266	1.428	ug/l #	94
20) Methylene Chloride	3.82	84	34764	2.526	ug/l	87
25) 2-Butanone	6.05	43	33928	27.270	ug/l #	85
39) Methylcyclohexane	8.84	83	514506	66.238	ug/l #	39
83) tert-Butylbenzene	14.19	119	47998	15.378	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061220.D
Acq On : 20 Mar 2019 20:48
Operator : VA/AP
Sample : K2004-07
Misc : 4.39g/5ml/MSVOA D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-19B(7-8)

Quant Time: Mar 22 03:29:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

