

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061228.D
 Acq On : 21 Mar 2019 2:20
 Operator : VA/AP
 Sample : K1688-23
 Misc : 6.31g/5ml/MSVOA D/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JC-02-031919-B

Quant Time: Mar 22 02:06:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-7.04
34) 1,4-Difluorobenzene	8.22	114	207	50.00	ug/l	0.00
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-12.36
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-14.51

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		0.00	65	0	0.00	ug/l	
Spiked Amount	50.000			Recovery	=	0.00%	
35) Dibromofluoromethane		0.00	113	0	0.00	ug/l	
Spiked Amount	50.000			Recovery	=	0.00%	
50) Toluene-d8		10.41	98	177	45.88	ug/l	0.01
Spiked Amount	50.000			Recovery	=	91.76%	
62) 4-Bromofluorobenzene		0.00	95	0	0.00	ug/l	
Spiked Amount	50.000			Recovery	=	0.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061228.D
Acq On : 21 Mar 2019 2:20
Operator : VA/AP
Sample : K1688-23
Misc : 6.31g/5ml/MSVOA D/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JC-02-031919-B

Quant Time: Mar 22 02:06:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

