

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061229.D
 Acq On : 21 Mar 2019 2:48
 Operator : VA/AP
 Sample : K1688-25
 Misc : 5.92g/5ml/MSVOA D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JC-02-031919-C

Quant Time: Mar 22 03:47:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	694380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	1173425	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	945660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	377772	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	275753	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
35) Dibromofluoromethane	6.92	113	442580	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
50) Toluene-d8	10.41	98	932897	42.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.32%	
62) 4-Bromofluorobenzene	13.55	95	364827	42.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.98%	
Target Compounds						
16) Acetone	3.23	43	15506	9.382	ug/l #	89
20) Methylene Chloride	3.82	84	54545	2.839	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061229.D
Acq On : 21 Mar 2019 2:48
Operator : VA/AP
Sample : K1688-25
Misc : 5.92g/5ml/MSVOA D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JC-02-031919-C

Quant Time: Mar 22 03:47:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

