

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061232.D
 Acq On : 21 Mar 2019 4:12
 Operator : VA/AP
 Sample : K1688-31
 Misc : 7.09g/5ml/MSVOA D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JC-02-031919-F

Quant Time: Mar 22 03:53:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	529943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	932168	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	654535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	227797	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	193322	43.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.66%	
35) Dibromofluoromethane	6.92	113	337898	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	10.40	98	698864	40.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.46%	
62) 4-Bromofluorobenzene	13.55	95	213697	31.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.40%	
Target Compounds						
16) Acetone	3.22	43	10527	8.345	ug/l	90
20) Methylene Chloride	3.81	84	41795	2.871	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061232.D
Acq On : 21 Mar 2019 4:12
Operator : VA/AP
Sample : K1688-31
Misc : 7.09g/5ml/MSVOA D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JC-02-031919-F

Quant Time: Mar 22 03:53:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

