

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061233.D
 Acq On : 21 Mar 2019 4:39
 Operator : VA/AP
 Sample : K1688-33
 Misc : 6.65g/5ml/MSVOA D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JC-02-031919-G

Quant Time: Mar 22 03:55:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	641020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1160983	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	812066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	352952	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	265147	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	6.92	113	404041	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
50) Toluene-d8	10.41	98	754657	34.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.76%	
62) 4-Bromofluorobenzene	13.55	95	292119	34.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.60%	
Target Compounds						
16) Acetone	3.22	43	39666	25.997	ug/l	97
20) Methylene Chloride	3.82	84	58938	4.200	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061233.D
Acq On : 21 Mar 2019 4:39
Operator : VA/AP
Sample : K1688-33
Misc : 6.65g/5ml/MSVOA D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JC-02-031919-G

Quant Time: Mar 22 03:55:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

