

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032019\
 Data File : VD061234.D
 Acq On : 21 Mar 2019 5:07
 Operator : VA/AP
 Sample : K1688-35
 Misc : 6.79g/5ml/MSVOA D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JC-02-031919-H

Quant Time: Mar 22 03:56:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	583822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	1116131	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	856780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	306079	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	246091	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	6.93	113	375405	45.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.64%	
50) Toluene-d8	10.40	98	898542	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	
62) 4-Bromofluorobenzene	13.55	95	300827	37.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.54%	
Target Compounds						
16) Acetone	3.23	43	33393	24.030	ug/l	94
20) Methylene Chloride	3.83	84	61550	5.570	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032019\
Data File : VD061234.D
Acq On : 21 Mar 2019 5:07
Operator : VA/AP
Sample : K1688-35
Misc : 6.79g/5ml/MSVOA D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JC-02-031919-H

Quant Time: Mar 22 03:56:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

