



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032924\
Data File : VD078788.D
Acq On : 29 Mar 2024 10:51
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Title : SW846 8260
Last Update : Fri Mar 15 05:38:03 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.9	3834	PASS
75	95	30	60	50.0	11349	PASS
95	95	100	100	100.0	22691	PASS
96	95	5	9	6.3	1435	PASS
173	174	0.00	2	1.4	307	PASS
174	95	50	100	94.1	21341	PASS
175	174	5	9	7.6	1624	PASS
176	174	95	101	98.0	20907	PASS
177	176	5	9	6.4	1346	PASS