

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041219\
 Data File : VD061779.D
 Acq On : 12 Apr 2019 15:37
 Operator : VA/AP
 Sample : K2200-05RE
 Misc : 6.65g/5ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-01-040919-CRE

Quant Time: Apr 12 17:18:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	495106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	796837	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	619008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	275215	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	238391	63.52	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	127.04%	
35) Dibromofluoromethane	6.91	113	372600	67.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	135.00%	
50) Toluene-d8	10.40	98	802240	57.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.68%	
62) 4-Bromofluorobenzene	13.55	95	327487	74.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	149.18%	
Target Compounds						
16) Acetone	3.22	43	13153	11.724	ug/l #	85
20) Methylene Chloride	3.81	84	67632	10.624	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041219\
Data File : VD061779.D
Acq On : 12 Apr 2019 15:37
Operator : VA/AP
Sample : K2200-05RE
Misc : 6.65g/5ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-01-040919-CRE

Quant Time: Apr 12 17:18:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

