

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041219\
 Data File : VD061789.D
 Acq On : 12 Apr 2019 20:23
 Operator : VA/AP
 Sample : K2360-02
 Misc : 5.14g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-BPM-01-E2-E2A-F2-F2A

Quant Time: Apr 12 23:56:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	548218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	897616	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	773165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	356293	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	221668	53.34	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	106.68%	
35) Dibromofluoromethane	6.92	113	334978	53.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.74%	
50) Toluene-d8	10.40	98	724449	46.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.74%	
62) 4-Bromofluorobenzene	13.55	95	289888	53.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.16%	
Target Compounds						
16) Acetone	3.21	43	27236	21.926	ug/l	96
20) Methylene Chloride	3.81	84	34777	4.934	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041219\
Data File : VD061789.D
Acq On : 12 Apr 2019 20:23
Operator : VA/AP
Sample : K2360-02
Misc : 5.14g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-BPM-01-E2-E2A-F2-F2A

Quant Time: Apr 12 23:56:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

