

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
 Data File : VD075679.D
 Acq On : 14 Apr 2023 14:57
 Operator : KP/SY
 Sample : 02307-01
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-02-041223

Quant Time: Apr 17 01:33:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	122426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	334523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	421685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	198907	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	78626	67.661	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.320%
35) Dibromofluoromethane	7.805	113	124496	55.332	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.660%
50) Toluene-d8	10.269	98	468776	52.027	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.060%
62) 4-Bromofluorobenzene	12.575	95	167280	69.848	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	139.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075679.D
Acq On : 14 Apr 2023 14:57
Operator : KP/SY
Sample : 02307-01
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-02-041223

Quant Time: Apr 17 01:33:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

