

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
 Data File : VD075681.D
 Acq On : 14 Apr 2023 15:58
 Operator : KP/SY
 Sample : 02308-01RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-01-041223RE

Quant Time: Apr 17 01:34:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.863	168	1881	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.775	114	4876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	2348	50.000	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	1405	78.692	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 157.380%	
35) Dibromofluoromethane	7.793	113	1922	58.605	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.200%	
50) Toluene-d8	10.263	98	7696	58.599	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 117.200%	
62) 4-Bromofluorobenzene	12.575	95	1993	57.093	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075681.D
Acq On : 14 Apr 2023 15:58
Operator : KP/SY
Sample : 02308-01RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-01-041223RE

Quant Time: Apr 17 01:34:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

