

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
 Data File : VD075686.D
 Acq On : 14 Apr 2023 18:32
 Operator : KP/SY
 Sample : 02262-09
 Misc : 11.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JPP-3A-040723

Quant Time: Apr 17 09:13:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	117440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	280231	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	309006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	93996	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	70511	63.254	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.500%
35) Dibromofluoromethane	7.805	113	105756	56.109	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.220%
50) Toluene-d8	10.269	98	380068	50.354	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.700%
62) 4-Bromofluorobenzene	12.575	95	100345	50.017	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075686.D
Acq On : 14 Apr 2023 18:32
Operator : KP/SY
Sample : 02262-09
Misc : 11.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JPP-3A-040723

Quant Time: Apr 17 09:13:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

