

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
 Data File : VD075689.D
 Acq On : 14 Apr 2023 20:04
 Operator : KP/SY
 Sample : 02347-02
 Misc : 4.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20230246-13-4

Quant Time: Apr 17 09:17:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	138243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	318368	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	530246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	217305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	77538	59.090	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.180%	
35) Dibromofluoromethane	7.804	113	110420	51.566	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.140%	
50) Toluene-d8	10.269	98	430179	50.166	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	12.575	95	217311	95.343	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 190.680%#	
Target Compounds						Qvalue
16) Acetone	4.028	43	6993	29.090	ug/l	93
17) Carbon Disulfide	4.269	76	145591	28.733	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075689.D
Acq On : 14 Apr 2023 20:04
Operator : KP/SY
Sample : 02347-02
Misc : 4.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20230246-13-4

Quant Time: Apr 17 09:17:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

