

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
 Data File : VD075690.D
 Acq On : 14 Apr 2023 20:35
 Operator : KP/SY
 Sample : 02347-03
 Misc : 4.18G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20230271-20-1

Quant Time: Apr 17 09:17:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	135616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	314418	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	411480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	210935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	71058	55.201	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.400%	
35) Dibromofluoromethane	7.805	113	104398	49.366	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.740%	
50) Toluene-d8	10.269	98	423733	50.035	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.060%	
62) 4-Bromofluorobenzene	12.575	95	173257	76.970	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 153.940%#	
Target Compounds						
						Qvalue
16) Acetone	4.022	43	22479	95.320	ug/l #	78
17) Carbon Disulfide	4.264	76	15580	3.134	ug/l	99
25) 2-Butanone	7.081	43	5346	17.607	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075690.D
Acq On : 14 Apr 2023 20:35
Operator : KP/SY
Sample : 02347-03
Misc : 4.18G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20230271-20-1

Quant Time: Apr 17 09:17:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

