



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041423\
Data File : VD075672.D
Acq On : 14 Apr 2023 09:15
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Title : SW846 8260
Last Update : Thu Apr 13 14:16:29 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.7	11069	PASS
75	95	30	60	48.3	34093	PASS
95	95	100	100	100.0	70515	PASS
96	95	5	9	6.0	4246	PASS
173	174	0.00	2	1.2	789	PASS
174	95	50	100	94.2	66397	PASS
175	174	5	9	7.3	4839	PASS
176	174	95	101	97.0	64395	PASS
177	176	5	9	6.6	4275	PASS