



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042123\
Data File : VD075767.D
Acq On : 21 Apr 2023 09:41
Operator : KP/SY
Sample : BFB117
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB117

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Title : SFAM01.0
Last Update : Wed Apr 26 02:39:47 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	11411	PASS
75	95	30	80	49.2	35851	PASS
95	95	100	100	100.0	72816	PASS
96	95	5	9	6.7	4877	PASS
173	174	0.00	2	1.2	810	PASS
174	95	50	120	95.3	69392	PASS
175	174	5	9	7.3	5085	PASS
176	174	95	105	95.2	66027	PASS
177	176	5	10	6.6	4333	PASS

Scan 1860 (12.528 min): VD075767.D\data.ms
BFB117

m/z	Abundance
40.00	831.0
44.00	306.0

Instrument :
MSVOA_D
ClientSampleId :
BFB117

m/z	Abundance
36.00	501.0
37.20	1993.0
38.10	2412.0
39.00	899.0
39.90	974.0
40.80	153.0
44.00	792.0
45.00	458.0
47.10	851.0
48.00	410.0
49.00	2782.0
50.10	11008.0
51.10	3243.0
52.00	186.0
56.00	1111.0
57.10	1945.0
59.90	622.0
61.00	3298.0
62.00	3230.0
63.10	2767.0
64.10	392.0
65.10	611.0
68.10	7062.0
69.00	5694.0
69.90	441.0
72.10	225.0
73.00	3309.0
74.00	12201.0
75.10	32272.0
76.00	2954.0
76.80	390.0
77.70	299.0
79.00	3883.0
79.90	721.0
80.90	3486.0
82.00	772.0
87.00	4219.0
88.00	3687.0
90.70	154.0
91.10	150.0
92.00	1937.0
93.00	2971.0
94.10	9702.0
95.00	65120.0
96.10	4391.0
97.00	170.0
104.00	343.0
106.10	256.0
107.80	166.0
115.90	244.0
116.90	702.0
117.90	374.0
119.00	311.0
127.80	284.0
129.90	433.0
132.90	313.0
136.80	163.0
140.90	594.0
143.00	1109.0
144.90	579.0
146.00	207.0
147.90	185.0
172.10	734.0
172.90	730.0
174.00	60624.0
175.00	4742.0
176.00	58640.0
177.00	4027.0
281.00	1500.0
282.00	376.0
283.10	276.0

Instrument :
MSVOA_D
ClientSampleId :
BFB117

m/z	Abundance
36.00	563.0
37.10	2700.0
38.00	2615.0
39.00	1365.0
40.00	1238.0
44.00	1021.0
45.00	524.0
47.20	680.0
47.90	446.0
49.10	3408.0
50.10	12838.0
51.00	3998.0
52.10	219.0
54.90	242.0
56.00	1087.0
57.00	1815.0
60.10	663.0
61.10	4003.0
62.10	3541.0
63.10	3492.0
64.10	458.0
65.00	593.0
68.00	8233.0
69.00	8567.0
70.00	532.0
71.90	437.0
73.10	3773.0
74.10	14214.0
75.10	40288.0
76.00	3475.0
77.00	676.0
78.00	342.0
78.90	3286.0
79.90	1003.0
81.00	3586.0
81.90	876.0
82.90	194.0
87.00	4537.0
88.00	3934.0
90.90	275.0
92.00	2484.0
93.10	3416.0
94.10	10276.0
95.10	81632.0
96.10	5952.0
103.90	342.0
105.80	293.0
115.90	367.0
116.90	454.0
117.80	401.0
118.90	737.0
124.90	273.0
127.80	321.0
129.80	230.0
131.00	166.0
132.90	220.0
133.80	208.0
134.70	176.0
136.70	297.0
140.90	668.0
142.90	960.0
145.10	388.0
147.80	247.0
155.10	234.0
170.80	164.0
171.90	713.0
173.00	924.0
174.00	76608.0
175.00	5262.0
176.00	73920.0
177.00	4666.0
191.10	219.0
193.00	317.0
206.80	234.0
265.10	179.0
281.10	2585.0
282.00	541.0
283.10	362.0

Instrument :
MSVOA_D
ClientSampleId :
BFB117

m/z	Abundance
35.90	407.0
37.00	2389.0
38.10	2378.0
38.90	885.0
40.00	741.0
44.00	1128.0
45.00	701.0
47.00	814.0
48.00	452.0
49.10	2791.0
50.10	10388.0
51.10	3505.0
52.00	213.0
52.90	175.0
55.00	302.0
56.00	1072.0
57.10	1904.0
59.80	578.0
61.10	2756.0
62.00	3120.0
63.10	2497.0
64.10	367.0
65.00	417.0
67.00	167.0
68.00	7908.0
69.10	6223.0
70.10	566.0
72.00	473.0
73.00	3886.0
74.10	11549.0
75.10	34992.0
76.10	3070.0
77.20	530.0
77.80	285.0
79.00	2962.0
79.90	718.0
81.00	3701.0
81.90	394.0
87.00	4313.0
88.00	4215.0
90.90	329.0
92.00	2253.0
93.00	3640.0
94.10	9445.0
95.10	71696.0
96.10	4288.0
103.80	462.0
104.80	189.0
105.90	443.0
115.00	252.0
115.80	250.0
117.00	626.0
117.90	389.0
118.90	781.0
128.00	203.0
130.00	327.0
133.00	327.0
133.60	153.0
134.90	249.0
136.90	174.0
141.00	824.0
142.90	1173.0
145.00	498.0
147.80	193.0
155.10	196.0
160.90	173.0
171.10	203.0
172.00	816.0
173.00	776.0
174.00	70944.0
175.00	5252.0
176.00	65520.0
176.90	4306.0
191.20	410.0
193.10	241.0
207.00	315.0
248.90	374.0
265.10	305.0
281.10	2223.0
282.10	810.0
283.20	523.0

Instrument :
MSVOA_D
ClientSampleId :
BFB117