

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042522\
 Data File : VD072702.D
 Acq On : 25 Apr 2022 10:30
 Operator : VA/SY
 Sample : VD0425SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0425SBL01

Quant Time: Apr 26 01:12:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	567967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	991531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	923725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	418844	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	291813	42.472	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.940%
35) Dibromofluoromethane	7.908	113	315514	40.946	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.900%
50) Toluene-d8	10.332	98	1186309	44.290	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.580%
62) 4-Bromofluorobenzene	12.620	95	420594	40.207	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042522\
Data File : VD072702.D
Acq On : 25 Apr 2022 10:30
Operator : VA/SY
Sample : VD0425SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0425SBL01

Quant Time: Apr 26 01:12:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

