

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042822\
 Data File : VD072723.D
 Acq On : 28 Apr 2022 14:27
 Operator : VA/SY
 Sample : N2602-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 12-15-SB1-BK-PRISON

Quant Time: Apr 29 01:22:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D042722S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 27 16:01:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	5816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	9276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	9670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	2956	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	4994	64.674	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.340%
35) Dibromofluoromethane	7.908	113	4416	71.487	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	142.980%
50) Toluene-d8	10.326	98	11680	52.447	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.900%
62) 4-Bromofluorobenzene	12.620	95	3405	40.540	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042822\
Data File : VD072723.D
Acq On : 28 Apr 2022 14:27
Operator : VA/SY
Sample : N2602-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
12-15-SB1-BK-PRISON

Quant Time: Apr 29 01:22:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D042722S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 27 16:01:06 2022
Response via : Initial Calibration

