



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042822\
 Data File : VD072718.D
 Acq On : 28 Apr 2022 09:30
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D042722S.M
 Title : SW846 8260
 Last Update : Wed Apr 27 16:01:06 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	47880	PASS
75	95	30	60	53.4	126520	PASS
95	95	100	100	100.0	237141	PASS
96	95	5	9	7.0	16681	PASS
173	174	0.00	2	1.6	2685	PASS
174	95	50	100	72.0	170667	PASS
175	174	5	9	8.7	14768	PASS
176	174	95	101	96.4	164565	PASS
177	176	5	9	7.0	11476	PASS