

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050322\
 Data File : VD072769.D
 Acq On : 03 May 2022 15:26
 Operator : VA/SY
 Sample : N2641-04
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-24

Quant Time: May 04 01:26:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	4096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	8277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	7999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	3050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	4537	85.331	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 170.660%#	
35) Dibromofluoromethane	7.902	113	3670	63.134	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 126.260%	
50) Toluene-d8	10.332	98	9802	47.008	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 94.020%	
62) 4-Bromofluorobenzene	12.626	95	3450	44.372	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 88.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050322\
Data File : VD072769.D
Acq On : 03 May 2022 15:26
Operator : VA/SY
Sample : N2641-04
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-24

Quant Time: May 04 01:26:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

