

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
 Data File : VD072784.D
 Acq On : 04 May 2022 13:04
 Operator : VA/SY
 Sample : N2641-04
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-24

Quant Time: May 05 01:36:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	255	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.855	114	437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	254	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.56
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.267	65	63	19.033	ug/l	-0.06
Spiked Amount	50.000	Range	50 - 163	Recovery	=	38.060%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.332	98	347	31.519	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	63.040%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.985	85	62	17.595	ug/l #	43
21) trans-1,2-Dichloroethene	5.656	96	76	21.552	ug/l #	1
23) Vinyl Acetate	6.397	43	80	13.261	ug/l #	75
25) 2-Butanone	6.926	43	200	207.370	ug/l #	52
27) cis-1,2-Dichloroethene	7.291	96	59	14.608	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
Data File : VD072784.D
Acq On : 04 May 2022 13:04
Operator : VA/SY
Sample : N2641-04
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-24

Quant Time: May 05 01:36:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

