

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
 Data File : VD072786.D
 Acq On : 04 May 2022 14:01
 Operator : VA/SY
 Sample : N2659-02
 Misc : 5.38G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 22L097-B

Quant Time: May 05 01:37:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	320731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	631211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	625683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	293002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	212064	50.936	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.880%
35) Dibromofluoromethane	7.908	113	122144	27.553	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	55.100%
50) Toluene-d8	10.332	98	776956	48.859	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.720%
62) 4-Bromofluorobenzene	12.620	95	284590	47.997	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.000%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.220	59	8544	23.719	ug/l #	83
16) Acetone	4.156	43	40161	39.816	ug/l	96
25) 2-Butanone	7.209	43	15128	12.471	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
Data File : VD072786.D
Acq On : 04 May 2022 14:01
Operator : VA/SY
Sample : N2659-02
Misc : 5.38G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
22L097-B

Quant Time: May 05 01:37:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

