

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
 Data File : VD072797.D
 Acq On : 04 May 2022 19:16
 Operator : VA/SY
 Sample : N2677-05
 Misc : 6.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2,7

Quant Time: May 05 01:39:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	37670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	59206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	66797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	29708	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	33423	68.352	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.700%
35) Dibromofluoromethane	7.909	113	26754	64.342	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.680%
50) Toluene-d8	10.332	98	74919	50.229	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.460%
62) 4-Bromofluorobenzene	12.620	95	28921	52.001	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050422\
Data File : VD072797.D
Acq On : 04 May 2022 19:16
Operator : VA/SY
Sample : N2677-05
Misc : 6.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2,7

Quant Time: May 05 01:39:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

