

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072806.D
 Acq On : 05 May 2022 13:21
 Operator : VA/SY
 Sample : N2677-05
 Misc : 7.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2,7

Quant Time: May 06 03:48:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	319008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	579601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	572146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	258982	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	203221	49.076	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.160%
35) Dibromofluoromethane	7.908	113	198384	48.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.480%
50) Toluene-d8	10.332	98	714113	48.906	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.820%
62) 4-Bromofluorobenzene	12.620	95	276791	50.838	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.680%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.973	56	9475	1.309	ug/l #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
Data File : VD072806.D
Acq On : 05 May 2022 13:21
Operator : VA/SY
Sample : N2677-05
Misc : 7.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2,7

Quant Time: May 06 03:48:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

