

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072810.D
 Acq On : 05 May 2022 15:16
 Operator : VA/SY
 Sample : N2672-03
 Misc : 6.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 03

Quant Time: May 06 03:49:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	359747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	643808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	646350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	292373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	213919	45.809	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.620%
35) Dibromofluoromethane	7.908	113	214910	47.530	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.060%
50) Toluene-d8	10.332	98	806073	49.699	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.620	95	312332	51.645	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.280%
Target Compounds						
					Qvalue	
30) Chloroform	7.708	83	21390	2.184	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
Data File : VD072810.D
Acq On : 05 May 2022 15:16
Operator : VA/SY
Sample : N2672-03
Misc : 6.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
03

Quant Time: May 06 03:49:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

