

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072817.D
 Acq On : 05 May 2022 18:37
 Operator : VA/SY
 Sample : N2672-10
 Misc : 5.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 10

Quant Time: May 06 03:51:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	441323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	797602	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	767734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	346714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	278629	48.637	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.280%		
35) Dibromofluoromethane	7.908	113	250861	44.783	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	89.560%		
50) Toluene-d8	10.332	98	992484	49.393	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	98.780%		
62) 4-Bromofluorobenzene	12.620	95	386279	51.556	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	103.120%		
Target Compounds						
					Qvalue	
30) Chloroform	7.703	83	25022	2.083	ug/l	89
68) m/p-Xylenes	11.838	106	15624	1.320	ug/l	95
69) o-Xylene	12.167	106	19066	1.725	ug/l	91
80) 1,3,5-Trimethylbenzene	12.944	105	58585	2.561	ug/l	99
84) 1,2,4-Trimethylbenzene	13.255	105	75789	3.315	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
Data File : VD072817.D
Acq On : 05 May 2022 18:37
Operator : VA/SY
Sample : N2672-10
Misc : 5.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
10

Quant Time: May 06 03:51:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

