

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072829.D
 Acq On : 06 May 2022 01:46
 Operator : VA/SY
 Sample : N2680-02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B1-5-7

Quant Time: May 06 03:55:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	161558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	292947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	301911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	136767	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	138828	66.198	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.400%
35) Dibromofluoromethane	7.908	113	109344	53.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.300%
50) Toluene-d8	10.332	98	358647	48.597	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.620	95	138765	50.426	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
Data File : VD072829.D
Acq On : 06 May 2022 01:46
Operator : VA/SY
Sample : N2680-02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B1-5-7

Quant Time: May 06 03:55:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

