

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072841.D
 Acq On : 06 May 2022 07:30
 Operator : VA/SY
 Sample : N2676-04
 Misc : 6.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-10-(23-25)

Quant Time: May 06 09:57:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	439044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	784116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	769103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	352023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	266762	46.807	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.620%
35) Dibromofluoromethane	7.908	113	250962	45.572	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.140%
50) Toluene-d8	10.332	98	1000202	50.633	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.620	95	387231	52.572	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.140%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.409	76	18340	1.102	ug/l #	90
31) Cyclohexane	7.979	56	17206	1.727	ug/l #	22
39) Methylcyclohexane	9.349	83	65930	6.348	ug/l	91
73) Isopropylbenzene	12.467	105	36294	1.325	ug/l	100
78) n-propylbenzene	12.808	91	64352	1.864	ug/l	99
85) sec-Butylbenzene	13.390	105	52708	1.770	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
Data File : VD072841.D
Acq On : 06 May 2022 07:30
Operator : VA/SY
Sample : N2676-04
Misc : 6.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-10-(23-25)

Quant Time: May 06 09:57:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

