



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050522\
 Data File : VD072822.D
 Acq On : 05 May 2022 21:01
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Title : SW846 8260
 Last Update : Tue May 03 08:33:49 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1806

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	34171	PASS
75	95	30	60	56.2	89203	PASS
95	95	100	100	100.0	158677	PASS
96	95	5	9	6.2	9834	PASS
173	174	0.00	2	1.2	1365	PASS
174	95	50	100	73.8	117083	PASS
175	174	5	9	7.4	8637	PASS
176	174	95	101	97.3	113875	PASS
177	176	5	9	7.0	7957	PASS