

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050824\
 Data File : VD078889.D
 Acq On : 08 May 2024 15:16
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: May 09 04:44:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 19:40:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	95297	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	234566	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	101846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	60174	56.888	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.780%
35) Dibromofluoromethane	7.810	113	84429	49.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.040%
50) Toluene-d8	10.269	98	308131	46.620	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.240%
62) 4-Bromofluorobenzene	12.575	95	78232	41.358	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050824\
Data File : VD078889.D
Acq On : 08 May 2024 15:16
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: May 09 04:44:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 19:40:18 2024
Response via : Initial Calibration

