

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051022\
 Data File : VD072909.D
 Acq On : 10 May 2022 12:44
 Operator : VA/SY
 Sample : N2717-01
 Misc : 4.46G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 24702

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/11/2022
 Supervised By : Mahesh Dadoda 05/11/2022

Quant Time: May 11 05:49:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	311016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	612337	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	538933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	156449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	201777	49.979	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	99.960%		
35) Dibromofluoromethane	7.908	113	205999	47.901	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	95.800%		
50) Toluene-d8	10.332	98	685383	44.429	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	88.860%		
62) 4-Bromofluorobenzene	12.620	95	217051	37.734	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	75.460%		
Target Compounds						
11) Tert butyl alcohol	5.197	59	17727m	50.749	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051022\
Data File : VD072909.D
Acq On : 10 May 2022 12:44
Operator : VA/SY
Sample : N2717-01
Misc : 4.46G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
24702

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/11/2022
Supervised By : Mahesh Dadoda 05/11/2022

Quant Time: May 11 05:49:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

