

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051022\
 Data File : VD072912.D
 Acq On : 10 May 2022 14:09
 Operator : VA/SY
 Sample : N2739-05
 Misc : 4.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124035

Quant Time: May 11 05:50:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	390528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	725224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	698513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	321815	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	235797	46.514	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.020%
35) Dibromofluoromethane	7.914	113	235804	46.297	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.600%
50) Toluene-d8	10.332	98	889252	48.672	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.620	95	338492	49.687	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051022\
Data File : VD072912.D
Acq On : 10 May 2022 14:09
Operator : VA/SY
Sample : N2739-05
Misc : 4.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124035

Quant Time: May 11 05:50:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

